
Scar repair ointment

【Product Name】 Scar Repair Ointment

【Model Specifications】 Scar repair ointment is available in 5g, 10g, 15g, 20g, 30g, 50g, and 100g tubes, with a tolerance of $\pm 9\%$. Special specifications are available upon request.

【Product Performance, Structure, and Composition】 Scar repair ointment consists of silicone gel, stearic acid, medical glycerin, purified water, an ointment base, and a soft tube. The ointment base consists of stearic acid, liquid paraffin, white petrolatum, lanolin, and triethanolamine. The soft tube consists of a body and a cap made of PET (polyester resin).

【Key Performance Indicators】 1. The cytotoxicity of the materials used in the product should not exceed Grade II. 2. The product should not exhibit delayed hypersensitivity reactions or intradermal irritation.

【Scope of Application】 Intended for the improvement and prevention of pathological scarring. Not for use on unhealed wounds.

【Contraindications】

1. Use with caution if you have allergies.
2. Do not use if you are pregnant or breastfeeding.
3. Use with caution in children.

【Instructions】 Cleanse the skin thoroughly, then apply the product.

【Precautions】

1. Do not use after the expiration date.
2. Carefully inspect the packaging before use. Do not use if it is damaged or contaminated.
3. Read the instructions carefully before use.



One-time use, destroyed after use



Do not use if the packaging is damaged

【Storage conditions and methods】 Please store in a room with relative humidity not exceeding 80%, in a ventilated, dry, light-proof place, and without corrosive gases.

【Registrant/Manufacturer】 Changde Yinuo Biopharmaceutical Co., Ltd.

【Registrant/Manufacturer Address/Manufacturing Address】 Fengkang Avenue, Xicheng Community, Dajinggang Town, Anxiang County, Changde City, Hunan Province (within Jinggang Park)

【After-Sales Service】 Changde Yinuo Biopharmaceutical Co., Ltd.

【Production License Number】 Hunan Food and Drug Administration Medical Device

Production Permit No. 20210068

[Registration Certificate Number/Product Technical Requirements Number]

Hunan Medical Device Registration Permit No. 20212141963

[Manufacture Date] See outer box for details

[Product Expiration Date] 2 years

[Product Instruction Manual Revision Date] April 25, 2021

Note: Products without the Changde Yinuo Biopharmaceutical Co., Ltd. traceability code are counterfeit and substandard products.